

SUPPORT PATIENT'S HEAD

- Place a pillow or rolled towel behind the patient's neck to reduce pressure on electrode 5 and 10 if patient is lying

PULL HAIR BACK

- Remove glasses and hair accessories
- Tuck hair behind ears

PLACE HEADBAND

- Close clasp centered on forehead, with cable oriented downwards
- Go to the RECORD page, enter patient information, and proceed to the ELECTRODE page
- Slide down to just above ears

ADDITIONAL INFORMATION, PRECAUTIONS, AND WARNINGS

24 Hour Customer Support: 1-800-633-0333

Medium-Large Size Head Circumference: 53.0 to 60.0 cm

Training QR Code

! Use outside of the prescriber head circumference range will significantly negatively impact device effectiveness and the reliability of the resulting data, which may lead to misdiagnosis.

! MR Unsafe. Remove headband from patient prior to performing Magnetic Resonance Imaging (MRI).

! Operating temperature range: 15°C to 30°C

! Non-adhesive material: rubber latex

! Dispose of contents/contents in accordance with federal, state, regional, local and/or hospital rules and regulations.

! Frequent checks for skin irritation should be performed when the device is used in infants and pediatric patients.

! Do not use on patients with a known history of skin allergies to parabens.

! Indications for Use:

The Ceribell EEG Headband is an electroencephalogram (EEG) electrode array intended for single patient use in the recording of EEGs. The EEG Headband is intended for prescription use in the home, healthcare facility, or clinical research environment.

PREPARE ELECTRODE 1

- Rotate ELECTRODE 1 back and forth to complete a full rotation
- Repeat for 7 rotations

DISPENSE GEL

- Return electrode to upright position
- Squeeze plunger firmly to dispense gel
- Repeat Step 5 and 6 on other electrodes

CHECK DEVICE

- To fix electrodes, repeat rotations then squeeze plunger to dispense more gel
- Only connect the headband to the Ceribell EEG Device.
- Repeat as needed until all electrodes on device appear green